

Approved by General Administration of Quality Supervision,



Inspection and Quarantine of the People's Republic of China

国家标准物质资源共享平台  
www.ncrm.org.cn  
GBW(E) 060140



国家标准物质资源共享平台  
www.ncrm.org.cn

## Certificate of Reference Material

Fenvalerate Purity



国家标准物质资源共享平台  
www.ncrm.org.cn

Sample Number:



国家标准物质资源共享平台  
www.ncrm.org.cn

Date of Certification:

National Institute of Metrology (NIM)

No.18,Bei San Huan Dong Lu, Chaoyang Dist, Beijing, 100013, P.R.China

Tel: +86(10) 64271639, 64278838, 64204432

Fax: +86(10) 64271639, 64213149

Website: [www.nim.ac.cn](http://www.nim.ac.cn) ; [www.ncrm.org.cn](http://www.ncrm.org.cn)



国家标准物质资源共享平台  
www.ncrm.org.cn



国家标准物质资源共享平台  
www.ncrm.org.cn

## 1. Description of Material & Intended Use

This certified reference material (CRM) is mainly used for pesticide residue test in areas such as environmental and food analysis, and for quality control, calibration of instruments, assessment of measurement methods and the like in pesticide production and test laboratory.

## 2. Sample Preparation

The CRM is produced from Fenvalerate through recrystallization (CAS No. : 51630-58-1).

## 3. Traceability and Certification

The traceability linkage is established using measuring protocols and instruments which meet the needs of the metrological community to be served. Two accurate and reliable measurement methods with distinct principles, GC (FID) and potential Titration, are used to determine and certify the purity of RM.

## 4. Property Values and Uncertainty

Number	Name	Certified value (%)	Relative expanded uncertainty (%) ( $k=2$ )
GBW (E) 060140	CRM of Fenvalerate pesticide	100.0	0.2

Certified purity of the main component and impurities such as moisture contributes to

the uncertainty of certified values of CRM.

国家标准物质资源共享平台

www.ncrm.org.cn

## 5. Constituency Check and Stability Test

According to the requirement of national technique criterion on CRM, the CRM aliquots are sampled, and consistency check and stability test in terms of purity are performed via GC (FID). The results demonstrated good consistency and stability.

The recommended minimum sampling amount is 8-15 mg.

This CRM shows good stability in follow-up test, and the period of validity of this CRM is set for 2 years from the Date of Certification. The stability of this CRM will be monitored by NIM in the future. The user will be promptly informed, if the value is found to vary during the period of validity.

## 6. Instructions for Package, Storage and Use

1. Package: This CRM is packaged in a brown glass bottle with screw-top, and each bottle contains no less than 0.2 g of this CRM.

2. Storage and Usage: This CRM should be stored in dark and dry condition at room temperature.

3. This RM is toxic and harmful. Appropriate protection is required when using it. The operator shall wear a face mask and latex gloves in order to avoid inhalation or skin contact.

国家标准物质资源共享平台  
www.ncrm.org.cn

After a small aliquot of the CRM is taken out, the bottle should be promptly sealed again and stored under the specified condition.



1. This CRM has a limited use in scientific research and analytical measurement. NIM is not responsible for any loss caused by improper use and storage of this CRM by the customer.
2. Please check the status of reference materials as soon as the sample arrives. The incurred maximum compensation will be exclusively restricted to the CRM itself, and no others should be claimed.
3. The certificate is only valid with the complete file stamped by NIM RM. Please keep the integrality of the certificate.
4. If more information about the CRM is required, please contact the technical service section.



National Institute of Metrology  
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China, 100013  
Tel: +86-10-64278838、84254143、64204432  
Fax: +86-10-64221811、64213149  
website: [www.nim.ac.cn](http://www.nim.ac.cn) ; [www.ncrm.org.cn](http://www.ncrm.org.cn)

